

CERTIFICATE OF CONFORMITY



1. HAZARDOUS LOCATION ELECTRICAL EQUIPMENT PER CANADIAN REQUIREMENTS

2. Certificate No: FM25CA0017X
3. Equipment:
(Type Reference and Name) UPT862 Ultra High Purity Pressure Transmitter
4. Name of Listing Company: WISE Control Inc
5. Address of Listing Company: 2022 Deogyong-Daero, Giheung-Gu,
Yongin-Si, Gyeonggi-do, 17097 Republic of Korea
6. The examination and test results are recorded in confidential report number:
PR470633 dated 27 February 2025

7. FM Approvals LLC, certifies that the equipment described has been found to comply with the following Approval standards and other documents:

CSA C22.2 No. 213:2021, CSA C22.2 No. 60529:R2021, CSA C22.2 No. 61010-1:2019

8. If the sign 'X' is placed after the certificate number, it indicates that the equipment is subject to specific conditions of use specified in the schedule to this certificate.
9. This certificate relates to the design, examination and testing of the products specified herein. The FM Approvals surveillance audit program has further determined that the manufacturing processes and quality control procedures in place are satisfactory to manufacture the product as examined, tested and Approved.
10. Equipment Ratings:

Nonincendive for Class I, Division 2, Groups ABCD, T6...T4;IP65
T6, Ta = -20°C to +55°C, T5, Ta = -20°C to +70°C, T4, Ta = -20°C to +80°C

11. The marking of the equipment shall include:

Class I, Division 2, Groups A, B, C, D; T6...T4;IP65
T6, Ta = -20°C to +55°C, T5, Ta = -20°C to +70°C, T4, Ta = -20°C to +80°C

Certificate issued by:

J.E. Marquedant
VP, Manager - Electrical Systems

27 February 2025

Date

To verify the availability of the Approved product, please refer to www.approvalguide.com

THIS CERTIFICATE MAY ONLY BE REPRODUCED IN ITS ENTIRETY AND WITHOUT CHANGE

FM Approvals LLC, One Technology Way, Norwood, MA 02062 USA

T: +1 (1) 781 762 4300 F: +1 (1) 781 762 9375 E-mail: information@fmaprovals.com www.fmaprovals.com

F 348 (Jul 24)

