

TEST REPORT



한국산업기술시험원
Korea Testing Laboratory

Report No. : 15-022716-01

Page of Pages : (1) / (2)



1. Client

Name : WISE CONTROL Inc.

Address : 2022, Deogyong-daero, Giheung-gu, Yongin-si, Gyeonggi-do, Korea

Date of Receipt : 2015. 04. 21

2. Use of Report : For Quality Control

3. Test Sample

Description : Bimetal thermometer with electrical contact

Manufacturer : WISE CONTROL Inc.

Model Name : T751

Serial Number :

Remark :

4. Date of Test : 2015. 04. 28. ~ 2015. 04. 29.

5. Test Standard/Method : IEC 60529:2001 (KS C IEC 60529:2006)

6. Testing Environment : Temperature : (21.3 ± 2.0) °C , Humidity : (46 ± 2) % R.H.

7. Test Results : Refer to the attached results

Note : 1. The test results contained apply only to the test sample(s) supplied by the client
2. This test report shall not be reproduced in full or in part without approval of the KTL in advance.

Affirmation	Tested by	(Signature)	Technical Manager	(Signature)
	Name : Kim Chang-ho		Name : KIM JI-HOON	

The above test report is the accredited test result by Korea Laboratory Accreditation Scheme, which signed the ILAC-MRA.

2015. 04. 30

Korea Testing Laboratory

Accredited by KOLAS, Republic of KOREA



87, Digital-ro 26-gil, Guro-gu, Seoul, KOREA Tel. +82-2-860-1531 Fax. +82-2-860-1549

FP202-03-03

※ 위 마크는 추후 전자확인증 대조 프로그램에서 원본대조시 사용되는 2D코드입니다.